

FACT SHEET FOR HEALTHCARE PROVIDERS

Yale School of Public Health, Department of Epidemiology of Microbial Diseases
SalivaDirect

Coronavirus
Disease 2019
(COVID-19)

This Fact Sheet informs you of the significant known and potential risks and benefits of the emergency use of the SalivaDirect.

The SalivaDirect is authorized for use with saliva specimens collected in the presence of a trained observer (adult trained on how to collect saliva samples) from individuals who are either suspected of COVID-19 by their healthcare provider or from individuals without symptoms or other epidemiological reasons to suspect COVID-19 using supervised or unsupervised specimen collection.

All patients whose specimens are tested with this assay will receive the Fact Sheet for Patients: Yale School of Public Health, Department of Epidemiology of Microbial Diseases - SalivaDirect.

What are the symptoms of COVID-19?

Many patients with COVID-19 have developed fever and/or symptoms of acute respiratory illness (e.g., cough, dyspnea), although some individuals experience only mild symptoms or no symptoms at all. The current information available to characterize the spectrum of clinical illness associated with COVID-19 suggests that, when present, symptoms include cough, shortness of breath or dyspnea, fever, chills, myalgias, headache, sore throat, new loss of taste or smell, nausea or vomiting or diarrhea. Signs and symptoms may appear any time from 2 to 14 days after exposure to the virus, and the median time to symptom onset is approximately 5 days. For further information on the symptoms of COVID-19 please see the link provided in “Where can I go for updates and more information?” section.

Public health officials have identified cases of COVID-19 throughout the world, including the United States. Please check the CDC COVID-19 webpage (see link provided in “Where can I go for updates and more information?” section at the end of this document) or your local jurisdictions website for the most up to date information.

This test is to be performed only using saliva specimens collected from individuals who are either suspected of COVID-19 by their healthcare provider or from individuals without symptoms or other epidemiological reasons to suspect COVID-19 using supervised or unsupervised specimen collection.

What do I need to know about COVID-19 testing?

Current information on COVID-19 for healthcare providers is available at CDC’s webpage, *Information for Healthcare Professionals* (see links provided in “Where can I go for updates and more information?” section).

- The SalivaDirect can be used to test saliva specimens collected without preservatives in a sterile container, in the presence of a trained observer (adult trained on how to collect saliva samples) or saliva specimens that are self-collected by individuals 18 years of age or older.
- The SalivaDirect should be ordered for the detection of COVID-19 in individuals who are either suspected of COVID-19 by their healthcare provider or from individuals without symptoms or other epidemiological reasons to suspect COVID-19 using supervised or unsupervised saliva collection.
- The SalivaDirect can be used to test saliva specimens that are self-collected by individuals 18 years of age or older unsupervised at home, and dropped off at a collection site, using the SalivaDirect Unsupervised Collection Kit or unsupervised at home using the SalivaDirect At-home Collection Kit and mailed to a testing laboratory, when determined to be appropriate by a healthcare provider.
- The SalivaDirect can also be used to test pooled samples containing up to five individual saliva specimens that are collected without preservatives in a sterile container in the presence of a trained observer (adult trained on how to collect saliva samples) from individuals suspected of COVID-19 by their healthcare provider or self-collected using the SalivaDirect Unsupervised Collection Kit or SalivaDirect At-home Collection Kit.

Report Adverse events, including problems with test performance or results, to MedWatch by submitting the online FDA Form 3500 (<https://www.accessdata.fda.gov/scripts/medwatch/index.cfm?action=reporting.home>) or by calling **1-800-FDA-1088**

FACT SHEET FOR HEALTHCARE PROVIDERS

Yale School of Public Health, Department of Epidemiology of Microbial Diseases
SalivaDirect

Coronavirus
Disease 2019
(COVID-19)

- The SalivaDirect is only authorized for use at laboratories designated by the Yale School of Public Health, Department of Epidemiology of Microbial Diseases, that includes the Clinical Molecular Diagnostics Laboratory, Department of Pathology, Yale School of Medicine, located at 310 Cedar St., New Haven, CT 06510, that are also certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a and meet the requirements to perform high complexity tests.

Specimens should be collected with appropriate infection control precautions. Current guidance is available at the CDC's website (see links provided in "*Where can I go for updates and more information?*" section).

When collecting and handling specimens from individuals suspected of being infected with COVID-19, appropriate personal protective equipment should be used as outlined in the CDC *Interim Laboratory Biosafety Guidelines for Handling and Processing Specimens Associated with Coronavirus Disease 2019 (COVID-19)*. For additional information, refer to CDC *Interim Guidelines for Collecting, Handling, and Testing Clinical Specimens from Persons Under Investigation (PUIs) for Coronavirus Disease 2019 (COVID-19)* (see links provided in "*Where can I go for updates and more information?*" section).

What does it mean if the specimen tests positive for the virus that causes COVID-19?

A positive test result for COVID-19 indicates that RNA from SARS-CoV-2 was detected, and therefore the patient is infected with the virus and presumed to be contagious. Laboratory test results should always be considered in the context of clinical observations and epidemiological data (such as local prevalence rates and current outbreak/epicenter locations) in making a final diagnosis and patient management decisions. Patient management should be made by a healthcare provider and follow current CDC guidelines.

The SalivaDirect has been designed to minimize the likelihood of false positive test results. However, it is still possible that this test can give a false positive result, even when used in locations where the prevalence is below 5%. In the event of a false positive result, risks to

patients could include the following: a recommendation for isolation of the patient, monitoring of household or other close contacts for symptoms, patient isolation that might limit contact with family or friends and may increase contact with other potentially COVID-19 patients, limits in the ability to work, delayed diagnosis and treatment for the true infection causing the symptoms, unnecessary prescription of a treatment or therapy, or other unintended adverse effects.

All laboratories using this test must follow the standard testing and reporting guidelines according to their appropriate public health authorities.

What does it mean if the specimen tests negative for the virus that causes COVID-19?

A negative test result for this test means that SARS-CoV-2 RNA was not present in the specimen above the limit of detection. However, a negative result does not rule out COVID-19 and should not be used as the sole basis for treatment or patient management decisions. It is possible to test a person too early or too late during COVID-19 infection to make an accurate diagnosis via SalivaDirect.

Specimens with low viral loads may not be detected in sample pools due to the decreased sensitivity of pooled testing. Your interpretation of negative results should take into account clinical and epidemiological risk factors.

When diagnostic testing is negative, the possibility of a false negative result should be considered in the context of a patient's recent exposures and the presence of clinical signs and symptoms consistent with COVID-19. The possibility of a false negative result should especially be considered if the patient's recent exposures or clinical presentation indicate that COVID-19 is likely, and diagnostic tests for other causes of illness (e.g., other respiratory illness) are negative.

If COVID-19 is still suspected based on exposure history together with other clinical findings, re-testing with a new sample with a sensitive method or without pooling should be considered by healthcare providers in consultation with public health authorities. Additional

Report Adverse events, including problems with test performance or results, to MedWatch by submitting the online FDA Form 3500 (<https://www.accessdata.fda.gov/scripts/medwatch/index.cfm?action=reporting.home>) or by calling 1-800-FDA-1088

FACT SHEET FOR HEALTHCARE PROVIDERS

Yale School of Public Health, Department of Epidemiology of Microbial Diseases
SalivaDirect

Updated: October 29, 2021

Coronavirus
Disease 2019
(COVID-19)

testing may be helpful to ensure testing was not conducted too early.

If a negative result is obtained with a saliva specimen and COVID-19 is still suspected based on exposure history together with other clinical findings, testing an alternative specimen type should be considered by healthcare providers in consultation with public health authorities.

Risks to a patient of a false negative test result include: delayed or lack of supportive treatment, lack of monitoring of infected individuals and their household or other close contacts for symptoms resulting in increased risk of spread of COVID-19 within the community, or other unintended adverse events.

The performance of this test was established based on the evaluation of a limited number of clinical specimens. The clinical performance has not been established in all circulating variants but is anticipated to be reflective of the prevalent variants in circulation at the time and location of the clinical evaluation. Performance at the time of testing may vary depending on the variants circulating, including newly emerging strains of SARS-CoV-2 and their prevalence, which change over time.

What is an EUA?

The United States FDA has made this test available under an emergency access mechanism called an Emergency Use Authorization (EUA). The EUA is supported by the Secretary of Health and Human Service's (HHS's) declaration that circumstances exist to justify the emergency use of in vitro diagnostics (IVDs) for the detection and/or diagnosis of the virus that causes COVID-19.

An IVD made available under an EUA has not undergone the same type of review as an FDA-approved or cleared IVD. FDA may issue an EUA when certain criteria are met, which includes that there are no adequate, approved, available alternatives, and based on the totality of scientific evidence available, it is reasonable to believe that this IVD may be effective in diagnosing COVID-19.

The EUA for this test is in effect for the duration of the COVID-19 declaration justifying emergency use of IVDs, unless terminated or revoked (after which the test may no longer be used).

What are the approved available alternatives?

Any tests that have received full marketing status (e.g., cleared, approved), as opposed to an EUA, by FDA can be found by searching the medical device databases here: <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases>. A cleared or approved test should be used instead of a test made available under an EUA, when appropriate and available. FDA has issued EUAs for other tests that can be found at:

<https://www.fda.gov/emergency-preparedness-and-response/mcm-legal-regulatory-and-policy-framework/emergency-use-authorization>.

Report Adverse events, including problems with test performance or results, to MedWatch by submitting the online FDA Form 3500 (<https://www.accessdata.fda.gov/scripts/medwatch/index.cfm?action=reporting.home>) or by calling **1-800-FDA-1088**

FACT SHEET FOR HEALTHCARE PROVIDERS

Yale School of Public Health, Department of Epidemiology of Microbial Diseases
SalivaDirect

Updated: October 29, 2021

Coronavirus
Disease 2019
(COVID-19)

Where can I go for updates and more information?

CDC webpages:

General: <https://www.cdc.gov/coronavirus/2019-ncov/index.html>

Symptoms:

<https://www.cdc.gov/coronavirus/2019-ncov/symptoms-testing/symptoms.html>

Healthcare Professionals:

<https://www.cdc.gov/coronavirus/2019-nCoV/hcp/index.html>

Information for Laboratories:

<https://www.cdc.gov/coronavirus/2019-nCoV/lab/index.html>

Laboratory Biosafety: <https://www.cdc.gov/coronavirus/2019-nCoV/lab-biosafety-guidelines.html>

Isolation Precautions in Healthcare Settings:

<https://www.cdc.gov/infectioncontrol/guidelines/isolation/index.html>

Specimen Collection: <https://www.cdc.gov/coronavirus/2019-nCoV/guidelines-clinical-specimens.html>

Infection Control: <https://www.cdc.gov/coronavirus/2019-ncov/php/infection-control.html>

FDA webpages:

General: www.fda.gov/novelcoronavirus

EUAs:(includes links to patient fact sheet and manufacturer's instructions) <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/in-vitro-diagnostics-euas>

LABORATORY CONTACT:

Contact information for the laboratory performing the SalivaDirect assay authorized test must be provided to the Healthcare Provider in the test report or material/mechanism (e.g. email) that accompanies this Fact Sheet and the test results.

Report Adverse events, including problems with test performance or results, to MedWatch by submitting the online FDA Form 3500 (<https://www.accessdata.fda.gov/scripts/medwatch/index.cfm?action=reporting.home>) or by calling **1-800-FDA-1088**